

**ANNEX 6**

CODE ..... **GT2723**  
OVERALL DIMENSION . . . 1000x1050x1300 mm  
WEIGHT ..... 60 kg

**IMPACT TEST CSTB SHOCK LOURD****STANDARDS**

CSTB CAHIER 3778-V6, ANNEX 6

**DESCRIPTION**

Instrument for determining impact resistance by heavy shock for unglazed ceramic tiles according to CSTB CAHIER 3778-V3-2018 – ANNEX 6

**TECHNICAL SPECIFICATIONS**

- Steel structure epoxy powder painted

**EQUIPMENT****ANNEX 6:**

Instrument for determining the heavy shock for unglazed ceramic tiles.  
The test apparatus consists of:

- Sand Plan of proper dimension for a cement plate of 70x70 cm with thickness of 6 cm
- A vertical rod where you can regulate a support with magnet with a steel ball run on it
- A steel ball of 510 gr
- Steel chrome ball 510 g